

Orthopedic Physical Therapy

Policy: Electric Equipment Inspection

All electrical equipment utilized by OPT using 120 volt current will be inspected at least yearly by a person qualified to test and approve such equipment for safety. A copy of the documentation stating items tested, results, and dates will be kept on file at the AFI office.


AFI Medical Services (1-508-668-6500)

In case of equipment malfunction, the equipment will be immediately unplugged and removed from the treatment area or office space. The clinical supervisor, office manager, and owners will be notified and repair/re-inspection will be performed before the equipment is used again.

Written:
Approved:

Reviewed/Revised:
Signature:

Reviewed/Revised: 12/2018

Signature: 

Reviewed/Revised:
Signature: